



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 BY: [unclear] ON: 1/17/2025 11:51 AM

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 102455	NAME OF AGENCY INDEPENDENCE PD	DATE OF INSPECTION 1/14/2025
LOCATION OF INSTRUMENT (STREET AND CITY) 14609 E TROMAN RD, INDEPENDENCE		TIME OF INSPECTION 0221

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION

COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER **ILT-X/METER** LOT # **AG311602** EXP. DATE **4/26/2025**

SIMULATOR TEMPERATURE (34°C ± 0.2°C) _____ SIM. SN _____ SIM. NIST EXP DATE _____

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .100	TEST 2 .101	TEST 3 .101
--------------------	--------------------	--------------------

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME B. SCHMIDT
---------------	---------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 240084, 4/2/2026	TELEPHONE NUMBER 816-325-7300
--	---

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 102455
Version no: 532B

TEST RECORD 00431

Temp Date Time ^{s/} 210L

Air Blank:
01/04/25 02:21 .000
Calibration Check:
32 01/04/25 02:21 .100

Subject Name

TEST 1

Subject I.D.

B.SCHMIDU

Operator Name, I.D.

240084

Location

14609 E TRUMAN RD

AS IV Serial no: 102455
Version no: 532B

TEST RECORD 00432

Temp Date Time ^{s/} 210L

Air Blank:
01/04/25 02:23 .000
Calibration Check:
32 01/04/25 02:23 .101

Subject Name

TEST 2

Subject I.D.

B.SCHMIDU

Operator Name, I.D.

240084

Location

14609 E TRUMAN RD

AS IV Serial no: 102455
Version no: 532B

TEST RECORD 00433

Temp Date Time ^{s/} 210L

Air Blank:
01/04/25 02:26 .000
Calibration Check:
32 01/04/25 02:26 .101

Subject Name

TEST 3

Subject I.D.

B.SCHMIDU

Operator Name, I.D.

240084

Location

14609 E TRUMAN RD

AS IV Serial no: 102455
Version no: 532B

TEST RECORD 00434

Temp Date Time ^{s/} 210L

VOID: RFI
12 01/04/25 02:28

Subject Name

TEST 4

Subject I.D.

B.SCHMIDU

Operator Name, I.D.

240084

Location

14609 E TRUMAN RD



Airgas USA LLC (LAB)
 3600 Bernard Street
 St. Louis, Mo, 63103
 PH: (314) 633-3100
 Fax: (314) 633-7328

Certificate of Analysis

Test Date: 26-Apr-2023

Customer Name
 Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Lot # AG311602 Model 108

Exp Date 26-Apr-2025	Cyl. Type 108	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% BrAC (272 ppm)
-------------------------	------------------	----------------------------------	--

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010561	391.8 ppm	EB0010603	392.6 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.04 ppm
EB0010661	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	800.0 ppm	CC727493	390.0 ppm
CC727496	263.0 ppm	CC727498	150.0 ppm

Analytical Method: NDIR

Digitally signed by Quality Control
 responsibility and standard certification of analysis
 Location: Airgas USA LLC (LAB)
 Date: 2023.04.26 16:21

Approved for Release: *Red Marsala*
 Red Marsala

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
 ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

PERMIT
TYPE II
BRETT SCHMIDLI

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Paula J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (R8-10)

DATE 4/2/2024

NUMBER 240084

EXPIRES 4/2/2026

MO 590-0771 (8-10)

STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator SCHMIDLI, BRETT
 Permit No 240084
 Date Issued 4/2/2024 Date Expires 4/2/2026

