



AS IV Serial no: 108264  
Version no: 532B

TEST RECORD 00689

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
01/17/25 21:37 .000  
Calibration Check:  
22 01/17/25 21:37 .100

Subject Name

TEST 1

Subject I.D.

TEST 1

Operator Name, I.D.

R. KENDALL 240186

Location

315 S. HADSON ST

Buckner PD

AS IV Serial no: 108264  
Version no: 532B

TEST RECORD 00690

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
01/17/25 21:40 .000  
Calibration Check:  
22 01/17/25 21:40 .099

Subject Name

TEST 2

Subject I.D.

TEST 2

Operator Name, I.D.

R. KENDALL 240186

Location

315 S. HADSON ST

Buckner PD

AS IV Serial no: 108264  
Version no: 532B

TEST RECORD 00691

Temp Date Time <sup>g/</sup> 210L

Air Blank:  
01/17/25 21:41 .000  
Calibration Check:  
22 01/17/25 21:41 .100

Subject Name

TEST 3

Subject I.D.

TEST 3

Operator Name, I.D.

R. KENDALL 240186

Location

315 S. HADSON ST

Buckner PD

AS IV Serial no: 108264  
Version no: 532B

TEST RECORD 00692

Temp Date Time <sup>g/</sup> 210L

VOID: RFI  
12 01/17/25 21:43

Subject Name

RFI

Subject I.D.

RFI

Operator Name, I.D.

SGT. R. KENDALL 240186

Location

315 S. HADSON ST

Buckner PD



**Airgas USA LLC (LAB)**  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 18-Mar-2024

**Lot #** AG407603 **Model** 108

<b>Exp Date</b> 16-Mar-2026	<b>Cyl. Type</b> 108	<b>Component</b> Ethanol Nitrogen	<b>Certified Concentration</b> 0.100 ± 2% BrAC (272 ppm)
--------------------------------	-------------------------	---	---

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	799.4 ppm	CC727493	389.8 ppm
CC727496	253.4 ppm	CC727498	150.2 ppm

**Analytical Method:** NDIR

Digitally signed by:Quality Control  
 Reason:Dry gas standard certification of analysis  
 Location:Airgas USA LLC (Lab)  
 Date:03.22.2024 07:48

Approved for Release: \_\_\_\_\_

Yusef Woods

**ISO 17025:2017 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT  
TYPE II**

**RICHARD L. KENDALL**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 8/29/2024

NUMBER 240186

EXPIRES 8/29/2026

*Mike Mason*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Douglas J. Nielsen*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)

 STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator KENDALL, RICHARD  
Permit No 240186  
Date Issued 8/29/2024 Date Expires 8/29/2026

