



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

RECEIVED
 By Tracy Crews at 10:37 am, Mar 12, 2025

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 108271	NAME OF AGENCY MISSOURI STATE HIGHWAY PATROL	DATE OF INSPECTION 03/06/2025
LOCATION OF INSTRUMENT (STREET AND CITY) MSHP TROOP D 3131 E. KEARNEY ST. SPRINGFIELD, MO 65803		TIME OF INSPECTION 11:39 am

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER GUTH LABORATORIES INC. LOT # 24310 EXP. DATE 08/27/2026
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 33.99 SIM. SN MP2418 SIM. NIST EXP DATE 12/04/2025

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .100	TEST 2 ← .101	TEST 3 ← .102
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	(0-.04)	(.05-.09)	(.10-.14)	(.15-.19)	(OVER .19)
----------	---------	-----------	-----------	-----------	------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

THIS INSTRUMENT IS OPERATING IN ACCORDANCE TO MODHSS STANDARDS

INSPECTING OFFICER	
SIGNATURE #357	PRINT NAME ROBERT G. WATKINS
TYPE II PERMIT NUMBER/EXPIRATION DATE 240129 5/29/2026	TELEPHONE NUMBER (417) 895-6868

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 108271
Version no: 532B

TEST RECORD 00825

Temp Date Time 210L^{9/}

Air Blank:

03/06/25 11:39 .000

Calibration Check:

11 03/06/25 11:39 .100

Subject Name

MAINT. TEST #1

Subject I.D.

Operator Name, I.D.

WATKINS #240129

Location

MSHP TROOP D

3131 E. KENNEDY ST
SPARKS MO

AS IV Serial no: 108271
Version no: 532B

TEST RECORD 00826

Temp Date Time 210L^{9/}

Air Blank:

03/06/25 11:41 .000

Calibration Check:

13 03/06/25 11:41 .101

Subject Name

MAINT. TEST #2

Subject I.D.

Operator Name, I.D.

WATKINS #240129

Location

MSHP TROOP D

3131 E. KENNEDY ST
SPARKS MO

AS IV Serial no: 108271
Version no: 532B

TEST RECORD 00827

Temp Date Time 210L^{9/}

Air Blank:

03/06/25 11:47 .000

Calibration Check:

16 03/06/25 11:47 .102

Subject Name

MAINT. TEST #3

Subject I.D.

Operator Name, I.D.

WATKINS #240129

Location

MSHP TROOP D

3131 E. KENNEDY ST
SPARKS MO

AS IV Serial no: 108271
Version no: 532B

TEST RECORD 00828

Temp Date Time 210L^{9/}

VOID: RFI

12 03/06/25 11:49

Subject Name

MAINT. TEST RFI

Subject I.D.

Operator Name, I.D.

WATKINS #240129

Location

MSHP TROOP D

3131 E. KENNEDY ST
SPARKS MO



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 24310 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on August 28, 2024, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is August 27, 2026 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.
All balances are calibrated annually by an outside agency using NIST traceable weights.
Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
ROBERT G. WATKINS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 5/29/2024

NUMBER 240129

EXPIRES 5/29/2026

Mike Massum

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David F. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator WATKINS, ROBERT
Permit No 240129
Date Issued 5/29/2024 **Date Expires** 5/29/2026

