



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN <u>111650</u>	NAME OF AGENCY <u>Bourbon PD</u>	DATE OF INSPECTION <u>1-2-25</u>
LOCATION OF INSTRUMENT (STREET AND CITY) <u>602 S. Franklin St Cuba</u>		TIME OF INSPECTION <u>2:25 pm</u>

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

<input checked="" type="checkbox"/> SIMULATOR SOLUTION	<input type="checkbox"/> COMPRESSED ETHANOL-GAS MIXTURE
<input checked="" type="checkbox"/> STANDARD SUPPLIER <u>Guth</u>	LOT # <u>23390</u> EXP. DATE <u>10-17-25</u>
<input checked="" type="checkbox"/> SIMULATOR TEMPERATURE (34°C ± 0.2°C) <u>34</u>	SIM. SN <u>MP3872</u> SIM. NIST EXP DATE <u>5-7-25</u>

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → <u>.100</u>	TEST 2 → <u>.100</u>	TEST 3 → <u>.100</u>
----------------------	----------------------	----------------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS <u>0</u>	(0-.04) <u>0</u>	(.05-.09) <u>0</u>	(.10-.14) <u>0</u>	(.15-.19) <u>0</u>	(OVER .19) <u>0</u>
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER	
SIGNATURE <u>Michael Centunzi</u>	PRINT NAME <u>Michael Centunzi</u>
TYPE II PERMIT NUMBER/EXPIRATION DATE <u>240089 / 4-8-26</u>	TELEPHONE NUMBER <u>573-732-4838</u>

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111650
Version no: 532B

TEST RECORD 00413

Temp Date Time 21:00

Air Blank: 01/02/25 14:26 .000

Calibration Check: 23 01/02/25 14:26 .100

Subject Name

Blank 1

Subject I.D.

Test 1

Operator Name, I.D.

Centurion / 240089

Location

602 S. Franklin St

Cuba

AS IV Serial no: 111650
Version no: 532B

TEST RECORD 00414

Temp Date Time 21:01

Air Blank: 01/02/25 14:29 .000

Calibration Check: 23 01/02/25 14:29 .100

Subject Name

Blank 2

Subject I.D.

Test 2

Operator Name, I.D.

Centurion / 240089

Location

602 S. Franklin St

Cuba

AS IV Serial no: 111650
Version no: 532B

TEST RECORD 00415

Temp Date Time 21:01

Air Blank: 01/02/25 14:31 .000

Calibration Check: 24 01/02/25 14:31 .100

Subject Name

Blank 3

Subject I.D.

Test 3

Operator Name, I.D.

Centurion / 240089

Location

602 S. Franklin St

Cuba

AS IV Serial no: 111650
Version no: 532B

TEST RECORD 00416

Temp Date Time 21:00

VOID: RFI 12 01/02/25 14:34

Subject Name

Blank 4

Subject I.D.

RPT

Operator Name, I.D.

Centurion / 240089

Location

602 S. Franklin St

Cuba



GUTH LABORATORIES, INC.

690 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **23390** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **October 18, 2023**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1207%** (w/vol) ethyl alcohol. The expiration date for this lot number is **October 17, 2025** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



Missouri Department of Health and Senior Services
 P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466

Paula Nickelson
 Acting Director

Mic

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP3872 Manufacturer: Guth
 Model Number: 12V500
 Agency: CUBA PD
 Agency Address: 602 S FRANKLIN, CUBA, MO 65453

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690 Bias: 0.00
 Uncertainty: 0.02
 Date of Certification: 10/27/2023 Date of Expiration: 10/27/2024

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.01	34.01	.02

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 5/7/2024
 Certification Expiration: 5/7/2025
 Simulator testing technician: M. BOND

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO
 Certification No: MP3872_572024

X *Brianna Medrano*

DHSS BAP Scientist Approving

Simulator Calibration Certification
 Issued by Lab Manager, DHSS BAP
 Revision Date: 06/25/2022

Breath Alcohol Program
 1903 Northwood Drive, Suite 4
 Poplar Bluff, MO 63901

DHSS BAP Document 3.6A
 Revision 2
 Page 1 of 1



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM



PERMIT
TYPE II

MICHAEL S. CENTUNZI

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and re and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 4/8/2024

NUMBER 240089

EXPIRES 4/8/2026

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Dave J. Nielsen

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (PS)

MO 680-0771 (6-10)

 STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator **CENTUNZI, MICHAEL**
Permit No **240089**
Date Issued **4/8/2024** Date Expires **4/8/2026**

