



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

| | | |
|-----------------------------|---|----------------------------------|
| ALCO SENSOR IV SN 111746 | NAME OF AGENCY Missouri State Highway Patrol | DATE OF INSPECTION 03/05/2025 |
|-----------------------------|---|----------------------------------|

| | |
|---|-----------------------------|
| LOCATION OF INSTRUMENT (STREET AND CITY) 3131 E Kearney, Springfield, Missouri 65803 | TIME OF INSPECTION 12:23 |
|---|-----------------------------|

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories, INC LOT # 24310 EXP. DATE 08/27/2026

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 33.99 SIM. SN MP2418 SIM. NIST EXP DATE 12/04/2025

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 → .098

TEST 2 → .099

TEST 3 → .099

RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

| | | | | | |
|----------|---------|-----------|--------------------|--------------------|------------|
| REFUSALS | (0-.04) | (.05-.09) | (.10-.14) <u>1</u> | (.15-.19) <u>2</u> | (OVER .19) |
|----------|---------|-----------|--------------------|--------------------|------------|

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

| | |
|---------------|---------------------------------------|
| SIGNATURE | PRINT NAME D W Henley # 727 |
|---------------|---------------------------------------|

| | |
|--|--|
| TYPE II PERMIT NUMBER/EXPIRATION DATE 240144 6/28/2026 | TELEPHONE NUMBER (417) 895 - 6868 |
|--|--|

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 111746
Version no: 532B

TEST RECORD 00699 s/
Temp Date Time 210L
Air Blank: 03/05/25 12:25 .000
Calibration Check: 18 03/05/25 12:25 .098

Subject Name
TEST 2
Subject I.D.

Operator Name, I.D.
D W HENLEY # 707
Location
3131 E KEARNEY

SPRINGFIELD, MO
DR # 707

AS IV Serial no: 111746
Version no: 532B

TEST RECORD 00700 s/
Temp Date Time 210L
Air Blank: 03/05/25 12:26 .000
Calibration Check: 18 03/05/25 12:26 .099

Subject Name
TEST 2
Subject I.D.

Operator Name, I.D.
D W HENLEY # 707
Location
3131 E KEARNEY

SPRINGFIELD, MO
DR # 707

AS IV Serial no: 111746
Version no: 532B

TEST RECORD 00701 s/
Temp Date Time 210L
Air Blank: 03/05/25 12:28 .000
Calibration Check: 19 03/05/25 12:28 .099

Subject Name
TEST 3
Subject I.D.

Operator Name, I.D.
D W HENLEY # 707
Location
3131 E KEARNEY

SPRINGFIELD, MO
DR # 707

AS IV Serial no: 111746
Version no: 532B

TEST RECORD 00702 s/
Temp Date Time 210L
VOID: RFI
12 03/05/25 12:29

Subject Name
RFI
Subject I.D.

Operator Name, I.D.
D W HENLEY # 707
Location
3131 E KEARNEY

SPRINGFIELD, MO
DR # 707



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4611 • TELEPHONE: 717-564-6470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 24310 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on August 28, 2024, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1209% (w/vol) ethyl alcohol. The expiration date for this lot number is August 27, 2026 at 11:59 PM.

When used in a calibrated Simulator, operating at 34°C +/- .2°C, this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

2

**PERMIT
 TYPE II**

DAVID W. HENLEY, JR.

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOX DMT

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 6/28/2024

NUMBER 240144

EXPIRES 6/28/2026

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David F. Nielson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (6-10)

LAB-4 (RS-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HENLEY, JR., DAVID

Permit No 240144

Date Issued 6/28/2024 Date Expires 6/28/2026

