



ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 111781	NAME OF AGENCY Warsaw Police Department	DATE OF INSPECTION 03/30/2025
-----------------------------	--	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 181 W. Harrison, Warsaw Mo 65355	TIME OF INSPECTION 5:49 pm
--	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)

TEMPERATURE OF ALCO SENSOR (10°C - 40°C)

PRINTER WORKING PROPERLY

TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE

STANDARD SUPPLIER Guth Laboratories LOT # 23390 EXP. DATE 10/17/2025

SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.02 SIM. SN MP6529 SIM. NIST EXP DATE 10/03/2025

CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ➡ .098	TEST 2 ➡ .098	TEST 3 ➡ .099
---------------	---------------	---------------

RFI DETECTOR OPERATING

**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

Instrument operating within established department of health limits

INSPECTING OFFICER

SIGNATURE 	PRINT NAME Sgt. Tony Helms
---------------	-------------------------------

TYPE II PERMIT NUMBER/EXPIRATION DATE 240099 05/07/2026	TELEPHONE NUMBER
--	------------------

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS TV Serial no: 111781
Version no: 532B

TEST RECORD 00254

Temp Date Time 2⁹/OL

Air Blank:
03/30/25 17:49 .000
Calibration Check:
24 03/30/25 17:49 .058

Subject Name

Test #1

Subject I.D.

Test #1

Operator Name, I.D.

T. Helms 111781

Location

181 W Harrison

Warsaw mo

AS TV Serial no: 111781
Version no: 532B

TEST RECORD 00255

Temp Date Time 2⁹/OL

Air Blank:
03/30/25 17:53 .000
Calibration Check:
24 03/30/25 17:53 .058

Subject Name

Test #2

Subject I.D.

Test #2

Operator Name, I.D.

T. Helms 111781

Location

181 W Harrison

Warsaw mo

AS TV Serial no: 111781
Version no: 532B

TEST RECORD 00256

Temp Date Time 2⁹/OL

Air Blank:
03/30/25 17:56 .000
Calibration Check:
25 03/30/25 17:56 .058

Subject Name

Test #3

Subject I.D.

Test #3

Operator Name, I.D.

T. Helms 111781

Location

181 W Harrison

Warsaw mo 65355

AS TV Serial no: 111781
Version no: 532B

TEST RECORD 00257

Temp Date Time 2⁹/OL

VOID: RFI
12 03/30/25 17:57

Subject Name

RFI Test

Subject I.D.

RFI Test

Operator Name, I.D.

T. Helms 111781

Location

181 W Harrison

Warsaw mo



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number 23390 of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on October 18, 2023, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain 0.1207% (w/vol) ethyl alcohol. The expiration date for this lot number is October 17, 2025 at 11:59 PM.

When used in a calibrated Simulator, operating at ~~34°C +/- 2°C,~~ this solution will give a breath alcohol analysis instrument reading of 0.100 g/210L +/- 3%.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



PERMIT
TYPE II
ANTHONY HELMS

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo. and 306.111 through 306.119 RSMo.

DATE 5/7/2024

NUMBER 240099

EXPIRES 5/7/2026

Mike Masom

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David J. Nielson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO-580-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator HELMS, ANTHONY
 Permit No 240099
 Date Issued 5/7/2024 Date Expires 5/7/2026

