



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
 STATE PUBLIC HEALTH LABORATORY
ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 119735	NAME OF AGENCY AURORA POLICE DEPARTMENT	DATE OF INSPECTION 01/02/2025
-----------------------------	--	----------------------------------

LOCATION OF INSTRUMENT (STREET AND CITY) 106 S. ELLIOTT, AURORA, MISSOURI, 65605.	TIME OF INSPECTION 6:29 am
--	-------------------------------

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

- DIGITAL READOUT (ALL ELEMENTS OPERATIONAL)
- TEMPERATURE OF ALCO SENSOR (10°C - 40°C)
- PRINTER WORKING PROPERLY
- TIME AND DATE DISPLAYING PROPERLY

BREATH ALCOHOL ACCURACY STANDARDS

- SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE
- STANDARD SUPPLIER GUTH LABS LOT # 24110 EXP. DATE 03/05/2026
- SIMULATOR TEMPERATURE (34°C ± 0.2°C) 34.00 SIM. SN SD2258 SIM. NIST EXP DATE 08/26/2025

- CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)
 Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)
 - 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE
 - 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE
 - 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 ← .097	TEST 2 ← .097	TEST 3 ← .096
---------------	---------------	---------------

- RFI DETECTOR OPERATING

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT: (DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS	0	(0-.04)	0	(.05-.09)	0	(.10-.14)	0	(.15-.19)	0	(OVER .19)	0
----------	---	---------	---	-----------	---	-----------	---	-----------	---	------------	---

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER

SIGNATURE 	PRINT NAME TATUM MAPLES
TYPE II PERMIT NUMBER/EXPIRATION DATE 240192 08/29/2026	TELEPHONE NUMBER (417) 678-5025

Return completed report to the: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

AS IV Serial no: 119735
Version no: 532C

TEST RECORD: 19092
Temp Date Time 9/
Air Blank: 01/02/25 06:37 000
Calibration Check: 19 01/02/25 06:37 097

Subject Name
maint. test
Subject I.D.
#1

Operator Name, I.D.
Tatum Myles #240192
Location
8/29/26
IOW. S. Elliott
Aurora, MO. 65605

AS IV Serial no: 119735
Version no: 532C

TEST RECORD: 00092
Temp Date Time 9/
Air Blank: 01/02/25 06:37 000
Calibration Check: 19 01/02/25 06:37 097

Subject Name
maint. test
Subject I.D.
#2

Operator Name, I.D.
Tatum Myles #240192
Location
8/29/26
IOW. S. Elliott
Aurora, MO. 65605

AS IV Serial no: 119735
Version no: 532C

TEST RECORD: 00093
Temp Date Time 9/
Air Blank: 01/02/25 06:41 000
Calibration Check: 21 01/02/25 06:41 095

Subject Name
maint. test
Subject I.D.
#3

Operator Name, I.D.
Tatum Myles #240192
Location
8/29/26
IOW. S. Elliott
Aurora, MO. 65605

AS IV Serial no: 119735
Version no: 532C

TEST RECORD: 00094
Temp Date Time 9/
VOID: RFI
12 01/02/25 06:44

Subject Name
RFI
Subject I.D.
RFI

Operator Name, I.D.
Tatum Myles #240192
Location
8/29/26
IOW. S. Elliott
Aurora, MO. 65605



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **24110** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography on **March 6, 2024**, using a Perkin Elmer Gas Chromatograph Autosystem XL S/N: 610N9030209, and found to contain **0.1215%** (w/vol) ethyl alcohol. The expiration date for this lot number is **March 5, 2026** at 11:59 PM.

When used in a calibrated Simulator, operating at $34^{\circ}\text{C} \pm .2^{\circ}\text{C}$, this solution will give a breath alcohol analysis instrument reading of **0.100 g/210L \pm 3%**.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

NIST Traceability:

Testing was conducted using Cerilliant Reference Standard lot number FN03072301 whose values are traceable to NIST.

All balances are calibrated annually by an outside agency using NIST traceable weights. Calibration verification is done prior to each use utilizing NIST traceable weights.