

**RECEIVED**

By Tracy Crews at 8:53 am, Mar 31, 2025

MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY**ALCO-SENSOR IV WITH PRINTER MAINTENANCE REPORT**

REPORT #7

Complete this report in duplicate at the time of the regular monthly preventative maintenance check, and whenever instrument is repaired. Send copy to Department of Health and Senior Services; retain original in department file.

ALCO SENSOR IV SN 120497	NAME OF AGENCY Buchanan County Sheriff's Office	DATE OF INSPECTION 03/28/2025
LOCATION OF INSTRUMENT (STREET AND CITY) 501 Faraon St, St. Joseph, MO 64501		TIME OF INSPECTION 7:54 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument. DIGITAL READOUT (ALL ELEMENTS OPERATIONAL) TEMPERATURE OF ALCO SENSOR (10°C - 40°C) PRINTER WORKING PROPERLY TIME AND DATE DISPLAYING PROPERLY**BREATH ALCOHOL ACCURACY STANDARDS** SIMULATOR SOLUTION COMPRESSED ETHANOL-GAS MIXTURE STANDARD SUPPLIER Intoximeters LOT # AG505603 EXP. DATE 02/25/2027 SIMULATOR TEMPERATURE (34°C ± 0.2°C) SIM. SN SIM. NIST EXP DATE CALIBRATION CHECK – (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE 0.080% STANDARD - MUST READ BETWEEN 0.076% and 0.084% INCLUSIVE 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

TEST 1 .098

TEST 2 .098

TEST 3 .098

 RFI DETECTOR OPERATING**INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)**

REFUSALS

(0-.04)

(.05-.09)

(.10-.14)

(.15-.19)

(OVER .19)

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

**INSPECTING OFFICER**

SIGNATURE

TYPE II PERMIT NUMBER/EXPIRATION DATE

250041 03/18/2027

PRINT NAME

Zackary Craft

TELEPHONE NUMBER

(816) 236-8800

**Return completed report to the:** Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office by mail, fax, or email.

# MONTHLY MAINTENANCE REPORT ASIV-W/PRINTER

AS IV Serial no: 120497  
Version no: 532C

TEST RECORD 00033  
Temp Date Time 210L g/

Air Blank: 03/28/25 07:54 .000  
Calibration Check: 22 03/28/25 07:54 .098

Subject Name  
Monthly Maintenance  
Subject I.D.

Operator Name, I.D.  
Craft 576  
Location

501 Faran St  
St. Joseph MO64504

AS IV Serial no: 120497  
Version no: 532C

TEST RECORD 00034  
Temp Date Time 210L g/

Air Blank: 03/28/25 07:56 .000  
Calibration Check: 23 03/28/25 07:56 .098

Subject Name  
Monthly Maintenance  
Subject I.D.

Operator Name, I.D.  
Craft 536  
Location

501 Faran St  
St. Joseph MO64504

AS IV Serial no: 120497  
Version no: 532C

TEST RECORD 00035  
Temp Date Time 210L g/

Air Blank: 03/28/25 07:57 .000  
Calibration Check: 23 03/28/25 07:57 .098

Subject Name  
Monthly Maintenance  
Subject I.D.

Operator Name, I.D.  
Craft 536  
Location

501 Faran St  
St. Joseph MO64504

AS IV Serial no: 120497  
Version no: 532C

TEST RECORD 00036  
Temp Date Time 210L g/

VOID: RFI  
12 03/28/25 07:59

Subject Name  
Monthly Maintenance  
Subject I.D.

Operator Name, I.D.  
Craft 536  
Location

501 Faran St  
St. Joseph MO



Airgas USA LLC (LAB)  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 25-Feb-2025

**Lot #** AG505603 **Model** 108

<b>Exp Date</b> 25-Feb-2027	<b>Cyl. Type</b> 108	<b>Component</b> Ethanol Nitrogen	<b>Certified Concentration</b> 0.100 ± 2% BrAC (272 ppm)
--------------------------------	-------------------------	---	---

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	799.4 ppm	CC727493	389.8 ppm
CC727496	253.4 ppm	CC727498	150.2 ppm

**Analytical Method:** NDIR

Digitally signed by:Quality Control  
 Reason:Dry gas standard certification of analysis  
 Location:Airgas USA LLC (Lab)  
 Date:02.28.2025 07:38

**Approved for Release:** \_\_\_\_\_  
 Yusef Woods

**ISO 17025:2017 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**ZACKARY CRAFT**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**ALCO-SENSOR IV WITH PRINTER, INTOX EC/IR II**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 3/18/2025

NUMBER 250041

EXPIRES 3/18/2027

*Adam J. Rubin*

DIRECTOR STATE PUBLIC HEALTH LABORATORY

*Sarah Willson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 680-0771 (6-10)

LAB-4 (R6-10)



STATE OF MISSOURI  
DEPARTMENT OF HEALTH AND SENIOR SERVICES  
BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

**Operator** CRAFT, ZACKARY

**Permit No** 250041

**Date Issued** 3/18/2025 **Date Expires** 3/18/2027

