

**RECEIVED**

By Tracy Crews at 3:17 pm, Jan 23, 2025



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM

**INTOX DMT MAINTENANCE REPORT**

REPORT #1

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 35 days).  
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

|  |  |   |
|--|--|---|
| INTOX DMT SN<br><b>500045</b>  | NAME OF AGENCY<br><b>Missouri State Highway Patrol</b> | DATE OF INSPECTION<br><b>01/15/2025</b> |
| LOCATION OF INSTRUMENT (STREET AND CITY)<br><b>2920 N. Shamrock Rd, Jefferson City</b> |  | TIME OF INSPECTION<br><b>08:29:22</b>   |

CHECKLIST: Place a mark in the box by each item if found to be satisfactory or is operating within established limits. (Write in observed values where determined). Unmarked items must be corrected before using instrument.

|   |  |
|---|--|
| <input checked="" type="checkbox"/> <b>DIAGNOSTIC RECORD</b>            |  |
| DATE AND TIME <u>01/15/2025 08:29:25</u>                                | <input checked="" type="checkbox"/> <b>DETECTOR</b>          |
| <input checked="" type="checkbox"/> <b>PROGRAM</b>                      | <input checked="" type="checkbox"/> <b>FILTER 1</b>          |
| <input checked="" type="checkbox"/> <b>SAMPLE CHAMBER</b> <u>48.8°C</u> | <input checked="" type="checkbox"/> <b>FILTER 2</b>          |
| <input checked="" type="checkbox"/> <b>BREATH TUBE</b> <u>48.1°C</u>    | <input checked="" type="checkbox"/> <b>FILTER 3</b>          |
| <input checked="" type="checkbox"/> <b>PUMP</b>                         | <input checked="" type="checkbox"/> <b>INTERNAL STANDARD</b> |

|  |   |
|--|---|
| <b>BREATH ANALYZER ACCURACY STANDARDS</b>          |   |
| <input type="checkbox"/> <b>SIMULATOR STANDARD</b> | <input checked="" type="checkbox"/> <b>COMPRESSED ETHANOL-GAS MIXTURE</b> |

|  |                              |                                    |
|--|------------------------------|------------------------------------|
| <input checked="" type="checkbox"/> <b>STANDARD SUPPLIER</b> <u>INTOXIMETERS</u> | <b>LOT #</b> <u>AG433904</u> | <b>EXP. DATE</b> <u>12/04/2026</u> |
| <input type="checkbox"/> <b>SIMULATOR TEMP</b> (34°C ± 0.2°C) _____              | <b>SIM. SN</b> _____         | <b>SIM. NIST EXP DATE</b> _____    |

|  |
|--|
| <input checked="" type="checkbox"/> <b>CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)</b><br>Run three tests using a standard. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard being used. |
| <input checked="" type="checkbox"/> <b>0.10% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE</b>  |
| <input type="checkbox"/> <b>0.08% STANDARD - MUST READ BETWEEN 0.076% AND 0.084% INCLUSIVE</b>   |
| <input type="checkbox"/> <b>0.04% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE</b>   |

|                             |                             |                             |
|-----------------------------|-----------------------------|-----------------------------|
| <b>TEST 1:</b> <u>0.099</u> | <b>TEST 2:</b> <u>0.099</u> | <b>TEST 3:</b> <u>0.098</u> |
|-----------------------------|-----------------------------|-----------------------------|

|  |
|--|
| <input checked="" type="checkbox"/> <b>PERFORM R.F.I. TEST</b> |
|--|

|   |          |            |            |            |             |
|---|----------|------------|------------|------------|-------------|
| <b>INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:</b> |          |            |            |            |             |
| <b>REFUSALS:</b> 0  | 0-.04: 0 | .05-.09: 0 | .10-.14: 0 | .15-.19: 0 | OVER .19: 0 |

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (USE OTHER SIDE IF NECESSARY)

Controller board replaced, standard change performed and Instrument calibrated.

|  |   |   |
|--|---|---|
| <b>INSPECTING OFFICER</b>              |   |   |
| SIGNATURE<br>                          | PRINT FULL NAME<br><b>JIMMY L CLEVELAND</b> |   |
| TYPE II PERMIT NUMBER<br><b>230082</b> | EXPIRATION DATE<br><b>05/03/2025</b>        | TELEPHONE NUMBER<br><b>573-751-4722</b> |

RETURN COMPLETED REPORT TO THE Breath Alcohol Program, Missouri Department of Health and Senior Services by mail, fax, or email



**Airgas USA LLC (LAB)**  
 3500 Bernard Street  
 St. Louis, Mo. 63103  
 Ph: (314) 533-3100  
 Fax: (314) 533-7328

## Certificate of Analysis

**Customer Name**  
 Exclusive Supplier  
 Intoximeters, Inc.  
 2081 Craig Road  
 St. Louis, Mo 63146

**Test Date:** 5-Dec-2024

**Lot #** AG433904 **Model** 108

|                 |                  |                     |                                |
|-----------------|------------------|---------------------|--------------------------------|
| <b>Exp Date</b> | <b>Cyl. Type</b> | <b>Component</b>    | <b>Certified Concentration</b> |
| 4-Dec-2026      | 108              | Ethanol<br>Nitrogen | 0.100 ± 2% BrAC (260 ppm)      |

**Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:**

| RGM Serial No. | Concentration | RGM Serial No. | Concentration |
|----------------|---------------|----------------|---------------|
| EB0010581      | 391.8 ppm     | EB0010603      | 392.5 ppm     |
| EB0010570      | 259.8 ppm     | EB0010559      | 258.9 ppm     |
| EB0010285      | 209.0 ppm     | EB0010562      | 104.2 ppm     |
| EB0010561      | 103.7 ppm     | EB0010579      | 52.94 ppm     |
| EB0010681      | 52.22 ppm     |                |               |

| CRM Serial No. | Concentration | CRM Serial No. | Concentration |
|----------------|---------------|----------------|---------------|
| CC727481       | 799.4 ppm     | CC727493       | 389.8 ppm     |
| CC727496       | 253.4 ppm     | CC727498       | 150.2 ppm     |

**Analytical Method:** NDIR

Digitally signed by:Quality Control  
 Reason:Dry gas standard certification of analysis  
 Location:Airgas USA LLC (Lab)  
 Date:12.06.2024 07:18

**Approved for Release:** \_\_\_\_\_  
 Yusef Woods

**ISO 17025:2017 A2LA accredited. Certificate Number 3082.06**  
**ISO 17034:2016 A2LA accredited. Certificate Number 3082.07**

STANDARD CHANGE

-----  
Missouri State Highway Patrol  
INTOX dmt: 500045  
-----

Date: 01/15/2025  
Time: 08:11:28

OPERATOR NAME:  
JIMMY L CLEVELAND  
PERMIT NUMBER: 230082  
EXPIRATION DATE: 05/03/2025


LOT #: AG433904  
SUPPLIER: INTOXIMETERS  
EXPIRATION: 12/04/2026  
SIMULATOR TYPE: DRY GAS

STANDARD INFORMATION  
CONCENTRATION: 0.100  
TARGET: 0.099

|                   |          |       |
|-------------------|----------|-------|
| BLANK TEST        | 0.000    | 08:12 |
| INTERNAL STANDARD | VERIFIED | 08:12 |
| EXTERNAL STANDARD | 0.098    | 08:12 |
| BLANK TEST        | 0.000    | 08:13 |

Average = 0.0980  
Std Dev = 0.0000  
Spread = 0.0000

*908001*



CALIBRATION FACTORS

-----  
Missouri State Highway Patrol  
INTOX dmt: 500045  
-----

Date: 12/17/2024  
Time: 08:19:21

OPERATOR NAME:  
JIMMY L CLEVELAND  
PERMIT NUMBER: 230082  
EXPIRATION DATE: 05/03/2025

LOT #: 23390  
SUPPLIER: GUTH  
EXPIRATION: 10/17/2025

Ca = 0.1000  
ADJ = 0.980970    0.800 <= ADJ < 1.200  
b1 = 0.0003        0.0000 <= b1 < 0.0040  
b2 = 0.0040        0.0010 <= b2 < 0.0100  
b3 = 0.0000        0.0000 <= b3 < 0.0040  
Xq = 0.1024        0.0500 <= Xq < 0.2500  
a21 = 1.125798    1.050 <= a21 < 1.300  
a31 = 0.453884    0.300 <= a31 < 0.800

*70811*



STATE OF MISSOURI  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM



**PERMIT**  
**TYPE II**  
**AARON S. CRAIG**

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

**INTOX DMT**

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 2/21/2024

NUMBER 240051

EXPIRES 2/21/2026

*Mike Massum*

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

*Paula J. Nickelson*

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

 **STATE OF MISSOURI**  
 DEPARTMENT OF HEALTH AND SENIOR SERVICES  
 BREATH ALCOHOL PROGRAM

**INSTRUMENT OPERATOR CARD**

*The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.*

Operator CRAIG, AARON  
 Permit No 240051  
 Date Issued 2/21/2024 Date Expires 2/21/2026

