

RECEIVED

By Tracy Crews at 7:48 am, Mar 10, 2025



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY

CMi INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-005852	LOCATION OF INSTRUMENT JOPLIN POLICE DEPT	DATE OF INSPECTION 03/05/2025	TIME OF INSPECTION 17:59
---------------------------------------	----------------------------------------------	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE	STANDARD LOT #	STANDARD EXPIRATION DATE
Air Blank	0.000	18:03	DRY	AG505005	02/19/2027
Cal Check	0.098	18:03	SIM TEMPERATURE	SIM SERIAL NUMBER	SIM CERTIFICATE EXPIRATION
Air Blank	0.000	18:04	N/A	N/A	N/A
Cal Check	0.098	18:04	STANDARD VALUE	STANDARD SUPPLIER	
Air Blank	0.000	18:04	0.100	INTOXIMETERS	
Cal Check	0.098	18:04	CALIBRATION CHECK RESULT 1		
Air Blank	0.000	18:04	0.098		
Cal Check	0.098	18:05	CALIBRATION CHECK RESULT 2		
Air Blank	0.000	18:05	0.098		
Cal Check	0.098	18:05	CALIBRATION CHECK RESULT 3		
Air Blank	0.000	18:05	0.098		
Pass			MAXIMUM DEVIATION (MUST BE WITHIN 6%)	SPREAD (MUST BE .005 OR LESS)	
			2.0%	0.000	

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass				
EEPROM Checksum Test	Pass		Air Blank	RFI*	18:06
Real Time Clock Test	Pass		Air Blank	0.000	18:06
DSP Test	Pass		*RFI Detect		
Analytical Stability Test	Pass		Pass		
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT						
REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19	
1	1	2	10	5	5	

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER		
SIGNATURE 	PRINT NAME KENNETH ANDERSEN	
TYPE II PERMIT NUMBER 230317	EXPIRATION DATE 12/21/2025	TELEPHONE NUMBER 4176233131



Airgas USA LLC (LAB)
 3500 Bernard Street
 St. Louis, Mo. 63103
 Ph: (314) 533-3100
 Fax: (314) 533-7328

Certificate of Analysis

Customer Name
Exclusive Supplier
 Intoximeters, Inc.
 2081 Craig Road
 St. Louis, Mo 63146

Test Date: 20-Feb-2025

Lot # AG505005 **Model** 55

Exp Date 19-Feb-2027	Cyl. Type 55	Component Ethanol Nitrogen	Certified Concentration 0.100 ± 2% BrAC (260 ppm)
--------------------------------	------------------------	-----------------------------------------	-------------------------------------------------------------

Certification Traceable to N.I.S.T. RGM and to CRM Ethanol Standards:

RGM Serial No.	Concentration	RGM Serial No.	Concentration
EB0010581	391.8 ppm	EB0010603	392.5 ppm
EB0010570	259.8 ppm	EB0010559	258.9 ppm
EB0010285	209.0 ppm	EB0010562	104.2 ppm
EB0010561	103.7 ppm	EB0010579	52.94 ppm
EB0010681	52.22 ppm		

CRM Serial No.	Concentration	CRM Serial No.	Concentration
CC727481	799.4 ppm	CC727493	389.8 ppm
CC727496	253.4 ppm	CC727498	150.2 ppm

Analytical Method: NDIR


Digitally signed by:Quality Control
 Reason:Dry gas standard certification of analysis
 Location:Airgas USA LLC (Lab)
 Date:02.20.2025 18:57



Approved for Release: _____

Yusef Woods

ISO 17025:2017 A2LA accredited. Certificate Number 3082.06
ISO 17034:2016 A2LA accredited. Certificate Number 3082.07



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM



**PERMIT
 TYPE II**

KENNETH ANDERSEN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/21/2023

NUMBER 230317

EXPIRES 12/21/2025

Mike Maxson

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

David J. Nicholson

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

MO 580-0771 (8-10)

LAB 4 (R6-10)



STATE OF MISSOURI
 DEPARTMENT OF HEALTH AND SENIOR SERVICES
 BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath form of expired air in Missouri.

Operator ANDERSEN, KENNETH

Permit No 230317

Date Issued 12/21/2023 Date Expires 12/21/2025

