



MISSOURI DEPARTMENT OF HEALTH AND SENIOR SERVICES
STATE PUBLIC HEALTH LABORATORY
CMI INTOXILYZER 8000 MAINTENANCE REPORT

REPORT #2

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send one copy to Department of Health and Senior Services, and retain one copy in department file.

INSTRUMENT SERIAL NUMBER 80-007514	LOCATION OF INSTRUMENT KANSAS CITY POLICE	DATE OF INSPECTION 01/03/2025	TIME OF INSPECTION 15:35
---------------------------------------	--	----------------------------------	-----------------------------

CALIBRATION CHECK RESULTS			CALIBRATION CHECK SUMMARY		
Test	g/210L	Time	STANDARD TYPE DRY	STANDARD LOT # 14323080A4	STANDARD EXPIRATION DATE 06/05/2025
Air Blank	0.000	15:37	SIM TEMPERATURE N/A	SIM SERIAL NUMBER N/A	SIM CERTIFICATE EXPIRATION N/A
Cal Check	0.082	15:37	STANDARD VALUE 0.080	STANDARD SUPPLIER CMI, INC.	
Air Blank	0.000	15:38	CALIBRATION CHECK RESULT 1 0.082		
Cal Check	0.083	15:38	CALIBRATION CHECK RESULT 2 0.083		
Air Blank	0.000	15:39	CALIBRATION CHECK RESULT 3 0.080		
Cal Check	0.080	15:39	MAXIMUM DEVIATION (MUST BE WITHIN 5%) 3.7%	SPREAD (MUST BE .005 OR LESS) 0.003	
Air Blank	0.000	15:40	Pass		

DIAGNOSTIC TEST RESULTS			RFI TEST RESULTS		
Voltage/Current Test	Pass		Test	g/210L	Time
RAM Test	Pass		Air Blank	0.000	15:40
EEPROM Checksum Test	Pass		Subject Test	RFI*	15:40
Real Time Clock Test	Pass		Air Blank	0.000	15:41
DSP Test	Pass		*RFI Detect		
Analytical Stability Test	Pass		Pass		
Modem Test	Pass				
Temperature Regulation Test	Pass				

NUMBER OF REFUSALS AND SUBJECT BREATH TESTS IN EACH RANGE SINCE LAST MAINTENANCE REPORT

REFUSALS	.00-.04	.05-.09	.10-.14	.15-.19	OVER .19
0	1	1	1	1	4

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary).

INSPECTING OFFICER		
SIGNATURE <i>PO Leonard J. Littleton #5626</i>	PRINT NAME PO LITTLETON #5626	
TYPE II PERMIT NUMBER 230323	EXPIRATION DATE 12/21/2025	TELEPHONE NUMBER 8162189393



STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

2



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com

PERMIT
TYPE II
JARED LITTLETON

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field service and repairs, and operate the following breath analyzer(s):

ALCO-SENSOR IV WITH PRINTER, INTOXILYZER 8000

for the determination of the alcoholic content of blood from a sample of expired air. Permit issued under the provisions of sections 577.020 through 577.041, RSMo and 306.111 through 306.119 RSMo.

DATE 12/21/2023

NUMBER 230323

EXPIRES 12/21/2025

MO 580-0771 (8-10)

Mike Mason

DIRECTOR OF STATE PUBLIC HEALTH LABORATORY

Daniel J. Nielsen

DIRECTOR OF DEPARTMENT OF HEALTH AND SENIOR SERVICES

LAB-4 (PB-10)

STATE OF MISSOURI
DEPARTMENT OF HEALTH AND SENIOR SERVICES
BREATH ALCOHOL PROGRAM

INSTRUMENT OPERATOR CARD

The named cardholder is authorized to operate an evidential breath alcohol instrument for the determination of the alcoholic content in breath from of expired air in Missouri.

Operator LITTLETON, JARED
Permit No 230323
Data Issued 12/21/2023 Date Expires 12/21/2025



Certificate of Analysis

Certificate ID: 15762
Part #: BAC105L080T
Cylinder Size: 105L
Lot Number: 14323080A4
Expiration: 6/5/2025

0.080 BAC (For the calibration of instruments used to determine breath alcohol concentration)

Contents: 105 Liters @ 1000 psig 70°F (21°C)

Component:	Reported Concentration:	Analytical Accuracy (U, k=2):	Analytical Method:	Distributed by:
Ethanol	288 ppm	+/- 0.002 BAC (G/210L) [5.2 ppm]	NDTR	CMI Inc. 316 East Ninth Street Owensboro, KY 42303 Phone 866-835-0690 www.alcoholtest.com
Nitrogen	Balance			

*Traceable to:
Certified Reference Material - 261.0 µmol/mol
Ethanol in Nitrogen - Serial No. ND7017 Lot No. 080722E2

Store in dry area, away from sources of heat, ignition and direct sunlight. Do not allow storage area to exceed 52 °C (125 °F).

[Signature]
Specialty Gas Lab Tech

06-01-2023
Issuance Date

[Signature]



The calibration results within this certificate were obtained at the facility listed above using equipment and standards capable of producing analytical results traceable to NIST, and apply only to the items consisted on this certificate. ILMO Products Company makes no warranty or representation as to the suitability of the use of any information provided for any particular purpose. The information use is at the sole discretion and risk of the user. Liability shall be limited to established replacement cost of this material or service. This certificate applies only to the items described and shall not be reproduced other than in full, without written approval from the issuing facility.

ISO/IEC 17025:2017 Accredited Laboratory