



**Missouri Department of Health and Senior Services**

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010  
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466

**Randall W. Williams, MD, FACOG**  
Director



**Michael L. Parson**  
Governor

# SIMULATOR CERTIFICATION REPORT

## SIMULATOR INFORMATION

**Simulator Serial Number:** MP2495      **Manufacturer:** Guth  
**Model Number:** 12V500  
**Agency:** MSHP (GHQ)  
**Agency Address:** 1510 E ELM ST, JEFFERSON CITY, MO 65101

## NIST THERMOMETER INFORMATION

**Serial Number:** 307715      **Bias:** 0.00  
**Uncertainty:** 0.02  
**Date of Certification:** 9/2/2019      **Date of Expiration:** 9/2/2020

## ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

## VERIFICATION RESULTS

| <u>Simulator Average</u> | <u>NIST Average</u> | <u>Combined Uncertainty</u> |
|--------------------------|---------------------|-----------------------------|
| 34.02                    | 34.03               | .03                         |

The combined uncertainty is calculated with a k=2 value.

## ADJUSTMENT RESULTS

No adjustment was needed.

**Date of testing:** 1/7/2020  
**Certification Expiration:** 1/7/2021  
**Simulator testing technician:** J. CLEVELAND

**Notes on Condition:** none

**Deviation(s) from method:** none

**DHSS BAP Scientist Approving:** B. LUTMER  
**Certification No:** MP2495\_172020

X

DHSS BAP Scientist Approving