



**Missouri Department of Health and Senior Services**

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010  
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466

**Randall W. Williams, MD, FACOG**  
Director



**Michael L. Parson**  
Governor

# **SIMULATOR CERTIFICATION REPORT**

## **SIMULATOR INFORMATION**

**Simulator Serial Number:** MP2109      **Manufacturer:** Guth  
**Model Number:** 12V500  
**Agency:** MISSOURI SAFETY CENTER  
**Agency Address:** 1200 SOUTH HOLDEN ST, WARRENSBURG, MO 64093

## **NIST THERMOMETER INFORMATION**

**Serial Number:** 17KMM00689      **Bias:** 0.00  
**Uncertainty:** 0.02  
**Date of Certification:** 12/15/2020      **Date of Expiration:** 12/15/2021

## **ENVIRONMENTAL CONDITIONS**

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

## **VERIFICATION RESULTS**

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	34.00	.02

The combined uncertainty is calculated with a k=2 value.

## **ADJUSTMENT RESULTS**

No adjustment was needed.

**Date of testing:** 1/19/2021  
**Certification Expiration:** 1/19/2022  
**Simulator testing technician:** M. BOND

**Notes on Condition:** none

**Deviation(s) from method:** none

**DHSS BAP Scientist Approving:** B. LUTMER  
**Certification No:** MP2109\_1192021

X

DHSS BAP Scientist Approving