



**Missouri Department of Health and Senior Services**

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010  
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466

**Donald G. Kauerauf**  
Director



**Michael L. Parson**  
Governor

# SIMULATOR CERTIFICATION REPORT

## SIMULATOR INFORMATION

**Simulator Serial Number:** MP6280      **Manufacturer:** Guth  
**Model Number:** 12V500  
**Agency:** DHSS  
**Agency Address:** 1903 NORTHWOOD DR STE 4, POPLAR BLUFF, MO 63901

## NIST THERMOMETER INFORMATION

**Serial Number:** 19BMM01307      **Bias:** 0.00  
**Uncertainty:** 0.02  
**Date of Certification:** 6/1/2020      **Date of Expiration:** 6/1/2021

## ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

## VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
33.99	34.00	.03

The combined uncertainty is calculated with a k=2 value.

## ADJUSTMENT RESULTS

No adjustment was needed.

**Date of testing:** 5/21/2021  
**Certification Expiration:** 5/21/2022  
**Simulator testing technician:** B. LUTMER

**Notes on Condition:** none

**Deviation(s) from method:** none

**DHSS BAP Scientist Approving:** B. LUTMER  
**Certification No:** MP6280\_5212021

X

---

DHSS BAP Scientist Approving