



# SIMULATOR CERTIFICATION REPORT

## SIMULATOR INFORMATION

**Simulator Serial Number:** SD2068      **Manufacturer:** Guth  
**Model Number:** 10-4D  
**Agency:** SHREWSBURY PD  
**Agency Address:** 4400 SHREWSBURY AVE, SHREWSBURY, MO 63119

## NIST THERMOMETER INFORMATION

**Serial Number:** 17KMM00690      **Bias:** 0.00  
**Uncertainty:** 0.02  
**Date of Certification:** 11/6/2020      **Date of Expiration:** 11/6/2021

## ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

## VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	34.01	.03

The combined uncertainty is calculated with a k=2 value.

## ADJUSTMENT RESULTS

No adjustment was needed.

**Date of testing:** 9/8/2021  
**Certification Expiration:** 9/8/2022  
**Simulator testing technician:** D. DEBOARD

**Notes on Condition:** none

**Deviation(s) from method:** none

**DHSS BAP Scientist Approving:** B. LUTMER  
**Certification No:** SD2068\_982021

X

---

DHSS BAP Scientist Approving