



Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466



Robert J. Knodell
Acting Director

Michael L. Parson
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: SD2254 **Manufacturer:** Guth
Model Number: 10-4D
Agency: HARRISONVILLE PD
Agency Address: 205 N LEXINGTON ST., HARRISONVILLE, MO 64701

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690 **Bias:** 0.00
Uncertainty: 0.02
Date of Certification: 11/6/2020 **Date of Expiration:** 11/6/2021

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	34.01	.03

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 7/27/2021
Certification Expiration: 7/27/2022
Simulator testing technician: D. DEBOARD

Notes on Condition: none
Deviation(s) from method: none

DHSS BAP Scientist Approving: B. LUTMER
Certification No: SD2254_7272021

X 

DHSS BAP Scientist Approving