



**Missouri Department of Health and Senior Services**

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010  
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466



**Robert J. Knodell**  
Acting Director

**Michael L. Parson**  
Governor

# SIMULATOR CERTIFICATION REPORT

## SIMULATOR INFORMATION

**Simulator Serial Number:** SD2773      **Manufacturer:** Guth  
**Model Number:** 10-4D  
**Agency:** ST LOUIS CO DEPT OF JUSTICE SVCS  
**Agency Address:** 100 S CENTRAL, CLAYTON, MO 63105

## NIST THERMOMETER INFORMATION

**Serial Number:** 17KMM00690      **Bias:** 0.00  
**Uncertainty:** 0.02  
**Date of Certification:** 11/6/2020      **Date of Expiration:** 11/6/2021

## ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

## VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	33.99	.03

The combined uncertainty is calculated with a k=2 value.

## ADJUSTMENT RESULTS

No adjustment was needed.

**Date of testing:** 4/27/2021  
**Certification Expiration:** 4/27/2022  
**Simulator testing technician:** M. BOND

**Notes on Condition:** none

**Deviation(s) from method:** none

**DHSS BAP Scientist Approving:** B. LUTMER  
**Certification No:** SD2773\_4272021

X 

DHSS BAP Scientist Approving