



Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466



Robert J. Knodell
Acting Director

Michael L. Parson
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: SD3000 **Manufacturer:** Guth
Model Number: 10-4D
Agency: WINFIELD PD
Agency Address: PO BOX 59, WINFIELD, MO 63389

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00689 **Bias:** 0.00
Uncertainty: 0.02
Date of Certification: 12/15/2020 **Date of Expiration:** 12/15/2021

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	33.98	.04

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 5/12/2021
Certification Expiration: 5/12/2022
Simulator testing technician: D. DEBOARD

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: B. LUTMER
Certification No: SD3000_5122021

X

DHSS BAP Scientist Approving