

Missouri Department of Health and Senior Services P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466



Michael L. Parson Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: SD3148

Donald G. Kauerauf

Director

Manufacturer: Guth

Model Number: 10-4D

Agency: LEE'S SUMMIT PD

Agency Address: 10 NE TUDOR, LEE'S SUMMIT, MO 64086

0.02

NIST THERMOMETER INFORMATION

Serial Number:

17KMM00690 Bias:

0.00

Uncertainty:

Date of Certification: 11/6/2020

Date of Expiration: 11/6/2021

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Simulator Average 34.00 NIST Average 34.00 Combined Uncertainty .02

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing:7/27/2021Certification Expiration:7/27/2022Simulator testing technician:D. DEBOARD

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: B. LUTMER

Certification No:

SD3148 7272021

Suan Later

Simulator Calibration Certification Issued by Lab Manager, DHSS BAP Revision Date: 05/16/2018 Breath Alcohol Program 1903 Northwood Drive, Suite 4 Poplar Bluff, MO 63901 DHSS BAP Document 3.6A Revision 1 Page 1 of 1 Х

DHSS BAP Scientist Approving