



SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP2102 Manufacturer: Guth
Model Number: 12V500
Agency: CARTHAGE PD
Agency Address: 310 W 4TH, CARTHAGE, MO 64836

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690 Bias: 0.01
Uncertainty: 0.02
Date of Certification: 11/10/2021 Date of Expiration: 11/10/2022

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Table with 3 columns: Simulator Average (34.00), NIST Average (34.01), Combined Uncertainty (.03)

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 4/27/2022
Certification Expiration: 4/27/2023
Simulator testing technician: R. SCHILDKNECHT

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO

Certification No: MP2102_4272022

X Brianna Medrano (signature)

DHSS BAP Scientist Approving