

Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466





Michael L. Parson Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP2417 Manufacturer: Guth

Model Number: 12V500

Agency: MSHP (GHQ)

Agency Address: 1510 E ELM ST, JEFFERSON CITY, MO 65101

NIST THERMOMETER INFORMATION

Serial Number: 306168 **Bias:** 0.00

Uncertainty: 0.02

Date of Certification: 11/2/2021 **Date of Expiration:** 11/2/2022

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Simulator Average NIST Average Combined Uncertainty

33.96 .01

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

Simulator Average NIST Average Combined Uncertainty

33.99 0.015

Date of testing: 7/1/2022 **Certification Expiration:** 7/1/2023

Simulator testing technician: J. CLEVELAND

Notes on Condition: CALIBRATED THERMISTOR 2. STATUS CODE 8 ON STARTUP

Deviation(s) from method:

DHSS BAP Scientist Approving: BRIANNA MEDRANO

Certification No: MP2417 712022

DHSS BAP Scientist Approving



Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466



Michael L. Parson Governor

Paula Nickelson
Acting Director