

Missouri Department of Health and Senior Services P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466 Paula Nickelson



Michael L. Parson Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP2468

Acting Director

Manufacturer: Guth

Model Number: 12V500

MSHP (GHQ) Agency:

Agency Address: 1510 E ELM ST, JEFFERSON CITY, MO 65101

NIST THERMOMETER INFORMATION

Serial Number:	306168	Bias: 0.	00
Uncertainty:	0.02		
Date of Certification:	11/2/2021	Date of Expiration	n: 11/2/2022

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Simulator Average	NIST Average	Combined Uncertainty
33.91	33.91	.01

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

Simulator Average	<u>e NIS</u>	T Average	Combined Uncertainty		
34.00		34.00	0.015		
Date of testing:	7/5/2022				
Certification Expiration:	7/5/2023				
Simulator testing technician:	J. CLEVELAND				
Notes on Condition: STATUS CODE 8. THERMISTOR 1 ADJUSTED					
Deviation(s) from method: none					

BRIANNA MEDRANO

DHSS BAP Scientist Approving:

Certification No:

MP2468 752022

Diaw Mehro

DHSS BAP Scientist Approving

Breath Alcohol Program 1903 Northwood Drive, Suite 4 Poplar Bluff, MO 63901



Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466

Paula Nickelson Acting Director Michael L. Parson Governor

