


Acting Director

#### Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010





Michael L. Parson Governor

# SIMULATOR CERTIFICATION REPORT

# SIMULATOR INFORMATION

Simulator Serial Number: MP2474 Manufacturer: Guth

Model Number: 12V500

MSHP (GHQ) Agency:

Agency Address: 1510 E ELM ST, JEFFERSON CITY, MO 65101

### NIST THERMOMETER INFORMATION

**Serial Number:** 306168 Bias: 0.00

**Uncertainty:** 0.02

**Date of Certification:** 11/2/2021 **Date of Expiration:** 11/2/2022

## ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

# VERIFICATION RESULTS

NIST Average Simulator Average **Combined Uncertainty** 

34.04 33.99 .10

The combined uncertainty is calculated with a k=2 value.

#### ADJUSTMENT RESULTS

**Simulator Average NIST Average Combined Uncertainty** 

34.00 33.99 0.025

Date of testing: 2/1/2022 **Certification Expiration:** 2/1/2023

Simulator testing technician: J. CLEVELAND

**Notes on Condition: CALIBRATED** 

Deviation(s) from method: none

**DHSS BAP Scientist Approving:** BRIANNA MEDRANO

**Certification No:** MP2474 212022

DHSS BAP Scientist Approving



#### Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010 RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466



Michael L. Parson Governor

Paula Nickelson
Acting Director