



Paula Nickelson
 Acting Director

Michael L. Parson
 Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP5544 **Manufacturer:** Guth
Model Number: 12V500
Agency: BRANSON WEST PD
Agency Address: PO BOX 2229, BRANSON WEST, MO 65737

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690 **Bias:** 0.00
Uncertainty: 0.02
Date of Certification: 10/24/2022 **Date of Expiration:** 10/24/2023

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.01	34.01	.02

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 11/21/2022
Certification Expiration: 11/21/2023
Simulator testing technician: R. SCHILDKNECHT

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO
Certification No: MP5544_11212022

X

DHSS BAP Scientist Approving