



Paula Nickelson
Acting Director

Michael L. Parson
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: SD2293 Manufacturer: Guth
Model Number: 10-4D
Agency: MANCHESTER PD
Agency Address: 200 HIGHLANDS BLVD. DR., MANCHESTER, MO 63011

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690 Bias: 0.01
Uncertainty: 0.02
Date of Certification: 11/10/2021 Date of Expiration: 11/10/2022

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Table with 3 columns: Simulator Average (34.00), NIST Average (34.01), Combined Uncertainty (.03)

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 7/12/2022
Certification Expiration: 7/12/2023
Simulator testing technician: M. BOND

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO

Certification No: SD2293\_7122022

X [Signature]

DHSS BAP Scientist Approving