



# SIMULATOR CERTIFICATION REPORT

## SIMULATOR INFORMATION

**Simulator Serial Number:** SD3322      **Manufacturer:** Guth  
**Model Number:** 10-4D  
**Agency:** SULLIVAN PD  
**Agency Address:** 106 PROGRESS DR., SULLIVAN, MO 63080

## NIST THERMOMETER INFORMATION

**Serial Number:** 16BMM00335      **Bias:** 0.00  
**Uncertainty:** 0.02  
**Date of Certification:** 5/12/2021      **Date of Expiration:** 5/12/2022

## ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

## VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	33.92	.16

The combined uncertainty is calculated with a k=2 value.

## ADJUSTMENT RESULTS

No adjustment was needed.

**Date of testing:** 1/26/2022  
**Certification Expiration:** 1/26/2023  
**Simulator testing technician:** B. MEDRANO

**Notes on Condition:** none

**Deviation(s) from method:** none

**DHSS BAP Scientist Approving:** BRIANNA MEDRANO

**Certification No:** SD3322\_1262022

X *Brianna Medrano*

DHSS BAP Scientist Approving



**Missouri Department of Health and Senior Services**

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010  
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466

**Donald G. Kauerauf**  
Director



**Michael L. Parson**  
Governor