



**Paula Nickelson**  
 Acting Director

**Michael L. Parson**  
 Governor

# SIMULATOR CERTIFICATION REPORT

## SIMULATOR INFORMATION

**Simulator Serial Number:** MP2925      **Manufacturer:** Guth  
**Model Number:** 12V500  
**Agency:** MISSOURI SAFETY CENTER  
**Agency Address:** 1200 SOUTH HOLDEN ST, WARRENSBURG, MO 64093

## NIST THERMOMETER INFORMATION

**Serial Number:** 17KMM00690      **Bias:** 0.00  
**Uncertainty:** 0.02  
**Date of Certification:** 10/24/2022      **Date of Expiration:** 10/24/2023

## ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

## VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
33.96	33.97	.03

The combined uncertainty is calculated with a k=2 value.

## ADJUSTMENT RESULTS

No adjustment was needed.

**Date of testing:** 1/9/2023  
**Certification Expiration:** 1/9/2024  
**Simulator testing technician:** R. SCHILDKNECHT

**Notes on Condition:** none

**Deviation(s) from method:** none

**DHSS BAP Scientist Approving:** BRIANNA MEDRANO  
**Certification No:** MP2925\_192023

**X**

DHSS BAP Scientist Approving