



**Paula Nickelson**  
 Acting Director

**Michael L. Parson**  
 Governor

# SIMULATOR CERTIFICATION REPORT

## SIMULATOR INFORMATION

**Simulator Serial Number:** SD2231      **Manufacturer:** Guth  
**Model Number:** 10-4D  
**Agency:** KNOB NOSTER PD  
**Agency Address:** 201 N STATE, KNOB NOSTER, MO 65336

## NIST THERMOMETER INFORMATION

**Serial Number:** 17KMM00689      **Bias:** 0.00  
**Uncertainty:** 0.02  
**Date of Certification:** 1/23/2023      **Date of Expiration:** 1/23/2024

## ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

## VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	33.99	.03

The combined uncertainty is calculated with a k=2 value.

## ADJUSTMENT RESULTS

No adjustment was needed.

**Date of testing:** 11/16/2023  
**Certification Expiration:** 11/16/2024  
**Simulator testing technician:** R. SCHILDKNECHT

**Notes on Condition:** none

**Deviation(s) from method:** none

**DHSS BAP Scientist Approving:** BRIANNA MEDRANO  
**Certification No:** SD2231\_11162023

**X**

DHSS BAP Scientist Approving



## Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010  
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466



---

**Paula Nickelson**  
Acting Director

**Michael L. Parson**  
Governor