



Paula Nickelson
Acting Director

Michael L. Parson
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: SD2245 Manufacturer: Guth
Model Number: 10-4D
Agency: SIKESTON DPS
Agency Address: 215 N NEW MADRID, SIKESTON, MO 63801

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690 Bias: 0.00
Uncertainty: 0.02
Date of Certification: 10/24/2022 Date of Expiration: 10/24/2023

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Table with 3 columns: Simulator Average (34.00), NIST Average (33.97), Combined Uncertainty (.06)

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 5/24/2023
Certification Expiration: 5/24/2024
Simulator testing technician: M. BOND

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO
Certification No: SD2245_5242023

X [Signature]

DHSS BAP Scientist Approving