



Paula Nickelson
Acting Director

Michael L. Parson
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: SD2262 Manufacturer: Guth
Model Number: 10-4D
Agency: WILLARD PD
Agency Address: 795 HUGHES RD., WILLARD, MO 65781

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00689 Bias: 0.00
Uncertainty: 0.02
Date of Certification: 1/23/2023 Date of Expiration: 1/23/2024

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Table with 3 columns: Simulator Average (34.00), NIST Average (34.00), Combined Uncertainty (.02)

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 6/14/2023
Certification Expiration: 6/14/2024
Simulator testing technician: M. BOND

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO
Certification No: SD2262_6142023

X [Signature]

DHSS BAP Scientist Approving