



Paula Nickelson
Director

Michael L. Parson
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP2148 Manufacturer: Guth
Model Number: 12V500
Agency: MSHP (GHQ)
Agency Address: 1510 E ELM ST, JEFFERSON CITY, MO 65101

NIST THERMOMETER INFORMATION

Serial Number: 307715 Bias: 0.00
Uncertainty: 0.02
Date of Certification: 9/17/2024 Date of Expiration: 9/17/2025

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Table with 3 columns: Simulator Average (34.03), NIST Average (34.00), Combined Uncertainty (.06)

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 11/6/2024
Certification Expiration: 11/6/2025
Simulator testing technician: J. CLEVELAND

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO
Certification No: MP2148_1162024

X [Signature]

DHSS BAP Scientist Approving