



Paula Nickelson
Acting Director

Michael L. Parson
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP2951 Manufacturer: Guth
Model Number: 12V500
Agency: GREENWOOD PD
Agency Address: 709 WEST MAIN ST., GREENWOOD, MO 64034

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690 Bias: 0.00
Uncertainty: 0.02
Date of Certification: 10/27/2023 Date of Expiration: 10/27/2024

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Table with 3 columns: Simulator Average (34.00), NIST Average (34.00), Combined Uncertainty (.02)

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 3/4/2024
Certification Expiration: 3/4/2025
Simulator testing technician: R. SCHILDKNECHT

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO
Certification No: MP2951_342024

X [Signature]

DHSS BAP Scientist Approving



Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466



Paula Nickelson
Acting Director

Michael L. Parson
Governor