



Paula Nickelson
Director

Michael L. Parson
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP4942 Manufacturer: Guth
Model Number: 12V500
Agency: DNR MO STATE PARK RANGER
Agency Address: PO BOX 176, JEFFERSON CITY, MO 65102

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00689 Bias: 0.00
Uncertainty: 0.02
Date of Certification: 5/13/2024 Date of Expiration: 5/13/2025

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Table with 3 columns: Simulator Average (34.01), NIST Average (34.02), Combined Uncertainty (.03)

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 10/22/2024
Certification Expiration: 10/22/2025
Simulator testing technician: R. SCHILDKNECHT

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO
Certification No: MP4942\_10222024

X [Signature]

DHSS BAP Scientist Approving