



Paula Nickelson
Director

Michael L. Parson
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP4948 Manufacturer: Guth
Model Number: 12V500
Agency: MINER PD
Agency Address: 103 STATE HWY H, MINER, MO 63801

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690 Bias: 0.00
Uncertainty: 0.02
Date of Certification: 10/21/2024 Date of Expiration: 10/21/2025

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Table with 3 columns: Simulator Average (33.99), NIST Average (34.02), Combined Uncertainty (.06)

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 12/12/2024
Certification Expiration: 12/12/2025
Simulator testing technician: R. SCHILDKNECHT

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO

Certification No: MP4948_12122024

X Brianna Medrano (signature)

DHSS BAP Scientist Approving