



Paula Nickelson
 Director

Michael L. Parson
 Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: MP4951 **Manufacturer:** Guth
Model Number: 12V500
Agency: DESOTO PD
Agency Address: #17 BOYD ST., DESOTO, MO 63020

NIST THERMOMETER INFORMATION

Serial Number: 20KMM02022 **Bias:** 0.00
Uncertainty: 0.02
Date of Certification: 8/15/2024 **Date of Expiration:** 8/15/2025

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

<u>Simulator Average</u>	<u>NIST Average</u>	<u>Combined Uncertainty</u>
34.00	34.02	.04

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 10/24/2024
Certification Expiration: 10/24/2025
Simulator testing technician: B. LUTMER

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO

Certification No: MP4951_10242024

X

DHSS BAP Scientist Approving