



Paula Nickelson
Acting Director

Michael L. Parson
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: SD2763 Manufacturer: Guth
Model Number: 10-4D
Agency: BROOKFIELD PD
Agency Address: 116 W BROOKS, BROOKFIELD, MO 64620

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690 Bias: 0.00
Uncertainty: 0.02
Date of Certification: 10/27/2023 Date of Expiration: 10/27/2024

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Table with 3 columns: Simulator Average (34.00), NIST Average (34.00), Combined Uncertainty (.02)

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 1/4/2024
Certification Expiration: 1/4/2025
Simulator testing technician: R. SCHILDKNECHT

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO

Certification No: SD2763_142024

X [Signature]

DHSS BAP Scientist Approving