



Sarah Wilson
Director

Mike Kehoe
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: SD1615 Manufacturer: Guth
Model Number: 10-4D
Agency: NORMANDY PD
Agency Address: 7700 NATURAL BRIDGE RAOD, NORMANDY, MO 63121

NIST THERMOMETER INFORMATION

Serial Number: 17KMM00690 Bias: 0.00
Uncertainty: 0.02
Date of Certification: 10/21/2024 Date of Expiration: 10/21/2025

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Table with 3 columns: Simulator Average (34.00), NIST Average (34.03), Combined Uncertainty (.06)

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 3/11/2025
Certification Expiration: 3/11/2026
Simulator testing technician: R. SCHILDKNECHT

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO

Certification No: SD1615_3112025

X [Signature]

DHSS BAP Scientist Approving