



Sarah Wilson
Acting Director

Mike Kehoe
Governor

SIMULATOR CERTIFICATION REPORT

SIMULATOR INFORMATION

Simulator Serial Number: SD3322 Manufacturer: Guth
Model Number: 10-4D
Agency: SULLIVAN PD
Agency Address: 106 PROGRESS DR., SULLIVAN, MO 63080

NIST THERMOMETER INFORMATION

Serial Number: 19BMM01307 Bias: 0.00
Uncertainty: 0.02
Date of Certification: 8/15/2024 Date of Expiration: 8/15/2025

ENVIRONMENTAL CONDITIONS

The environmental conditions during testing are within the tolerances of DHSS BAP method 3.

VERIFICATION RESULTS

Table with 3 columns: Simulator Average (34.00), NIST Average (34.04), Combined Uncertainty (.09)

The combined uncertainty is calculated with a k=2 value.

ADJUSTMENT RESULTS

No adjustment was needed.

Date of testing: 2/26/2025
Certification Expiration: 2/26/2026
Simulator testing technician: B. MEDRANO

Notes on Condition: none

Deviation(s) from method: none

DHSS BAP Scientist Approving: BRIANNA MEDRANO

Certification No: SD3322_2262025

X [Signature]

DHSS BAP Scientist Approving



Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466



Sarah Wilson
Acting Director

Mike Kehoe
Governor